NewCare V1 Bed

Quick reference sheet





Bed's components

A	Removable headboard	H	Backrest spindle
B	Backrest side rails	I	Integral wheels Ø15 cm
C	Legrest side rails	J	Braking pedal
D	HPL mattress platform	L	Lifting pedal
E	Sheet steel central section	M	Protection discs
F	Removable footboard	P	Trend / Reverse Trend handle
G	Legrest spindle		

Cleaning

Daily cleaning and disinfection with up to a 1000ppm concentration of chlorine solutions:

- Rails. Headboard. Footboard.

Cleaning and disinfection after a patient leaves or before a transfer:

- Mattress platform area. Chassis of the bed.

Thorough cleaning and disinfection every two months:

- Support and lifting area of the bed.

Maintenance

Always check the correct state of:

- Mechanical drives.
- Wheels.
- Braking mechanisms.

For more information, please consult the use and maintenance manual.

www.pardo.es
TECHNICAL SERVICE CENTRE
(+34) 976 300 033

sept 2024 | CódPCO741794_QuickSheet_NewCareV1_EN_2020_rev02

Mechanical functions

Sections movements

Unlock crank:

- Pull the lever bed outwards.
- Direct it to its working position.
- Push the lever to the correct position.

Articulating sections:

- Raise: Turn clockwise the left crank handle..
- Lower: Turn anticlockwise the left crank handle.



Unlock crank



Working position

Rise the sections



Lower the sections

Trendelenburg / Reverse Trendelenburg positions

Trendelenburg tilting:

- Press the handle placed on footboard.
- Tilt the footboard related to the headboard until the maximum inclination is reached.

Reverse Trendelenburg tilting:

- Press the handle placed on footboard.
- Tilt the headboard related to the until the footboard maximum inclination is reached.



Other functions

Side rails (1):

- Raising the rail: Pull the rail upwards until it is immobilised in the highest position.
- Lowering the rail: Pull the handle found underneath the rail outwards. The rail will automatically lower.

Centralised braking system (2):

- Braking the bed: Pedal in low position.
- Moving the bed: Pedal in intermediate position.
- Activating steering wheel: Pedal in raised position.



Raising side rails





Lowering side rails



Brake



Free movement



Steering wheel

Height adjustment

Raise mattress platform:

- To operate the rising movement is stepping several times the green pedal until the bed arrives to the desirable height.



Lower mattress platform:

- To operate the lowering movement is stepping red pedal.
- If the red pedal is kept depressed, the bed lowers to its lowest position.



Industrias Hidráulicas Pardo S.L. company has the following certifications:

IQNET LTD · AENOR GESTIÓN DE LA CALIDAD ISO 9001 ES-NA-0170013 · AENOR GESTIÓN CALIDAD GESTIÓN DEL SOFTWARE SGS ISO 13485 ES-NA-0170016 · AENOR GESTIÓN AMBIENTAL ISO 14001 ES-NA-0170182 · AENOR ECODISEÑO ISO 14006 ES-NA-0170012 · CE MD

SAFETY WARNINGS

- The maximum duration of a continuous manoeuvre should not exceed 2 minutes, as damage may be caused to the equipment.
- Before raising and lowering the moving parts of the mattress platform, with the other parts of the bed or accessories, it is essential to check that there are no obstacles between them.
- The Trendelenburg and reverse Trendelenburg movement should always be performed under medical supervision.

pardo

Industrias Hidráulicas Pardo, S.L. C/ Los Ángeles 5, (Polígono Centrovía) - 50198 La Muela, Zaragoza (Spain)